

# Replacement Lamps for HPLC Detectors

by Rebecca Wittrig, Ph.D., HPLC Product Marketing Manager

- Meet or exceed original manufacturer's performance.
- Simplify paperwork—order parts when you order columns and consumables.

In addition to the lamps listed here, we have an extensive range of HPLC replacement parts and accessories. For a full listing, visit our website, or request publication 59012. If you don't see what you need, please call us—we are constantly adding new parts to meet your needs.



Description	Model #	Similar to OEM part #	qty.	cat.#
<b>For Agilent HPLC Systems</b>				
Detector Lamp, 1090 DA, 1050 VW/DA/MWD	1090, 1050	79883-60002	ea.	25260
Lamp, DAD G1315A, G1365A	1100	2140-0590	ea.	25261
Lamp, VWD G1314A	1100	G1314-60100	ea.	25262
8453 Deuterium Lamp	—	2140-0605	ea.	25263
G1321 Fluorescence Detector Flash Lamp	—	2140-0600	ea.	25264
<b>For Beckman HPLC Systems</b>				
Deuterium Lamp	DU60, 62, 64, 65	596791	ea.	25454
<b>For Hitachi HPLC Systems</b>				
Deuterium Lamp, Prealigned	L4000, L4200, L4250, L7400	885-3570	ea.	25465
<b>For PerkinElmer Instruments</b>				
Deuterium Lamp	PE Lambda--2, 5, 7, 8, 10, 11, 12, 14, 15, 16, 17, 18, 19, 20, 40, 800, 900	B0160917	ea.	25436
Deuterium Lamp	PE 200/785A	N2920149	ea.	25431
<b>For Shimadzu HPLC Systems</b>				
Deuterium Lamp	SPD-6A	062-65056-02	ea.	25283
Deuterium Lamp	SPD-10A, 10AV	228-34016-02	ea.	25284
<b>For Waters™ Detectors</b>				
Xenon Lamp (w/o holder or mirror)	470	—	ea.	25404
Xenon Lamp	474	—	ea.	25405
Deuterium Lamp (UV/Vis)	480, 481	99499	ea.	25403
Deuterium Lamp (UV/Vis)	484	80357	ea.	25406
Deuterium Lamp (UV/Vis)	486	80678	ea.	25407
Deuterium Lamp	996, 2996	WAT052586	ea.	25408
Deuterium Lamp	2487	WAS081142	ea.	25409
Deuterium Lamp, long life (2000 hours)	486	—	ea.	25410

## Free Literature!



**Genuine Restek Replacement Parts for HPLC Systems**  
(lit. cat.# 59012\*)

Over time, detector lamps, check valves, pump piston seals, and other components wear out or become contaminated. Working with defective parts means poor chromatography and, possibly, shortened column lifetimes. This 4-page publication lists a wide selection of Restek parts for Agilent, Beckman, Hitachi, PerkinElmer, Shimadzu, and Waters instruments, to keep these systems running smoothly and chromatography sharp. Genuine Restek Replacement Parts equal or exceed the performance of original equipment components.



**Coming soon... HPLC Columns and Accessories**  
(lit. cat.# 59241B\*)

New, expanded 2004-2005 edition features columns and bulk materials, instrument parts, innovative tools and accessories, and many example chromatograms.

\*When requesting literature from outside the US, add "-INT" to the lit. cat.#.



**Lit. Cat. # 59051**  
©2004 Restek Corp.

**Restek Trademarks:** Alumaseal, Crossbond, EZ No-Vent, MegaMix, Press-Tight, Resprep, Rtx, SeCure, Silcosteel, Silcote, Siltek, Stabilwax, and the Restek logo.

**Other Trademarks:** Agilent (Agilent Technologies, Inc.), Teflon, Vespel (E.I. du Pont de Nemours & Co., Inc.), Trace (Thermo Finnigan).

**Restek** Phone: 800-356-1688 or 814-353-1300 • Fax: 814-353-1309 • www.restek.com  
110 Benner Circle, Bellefonte, PA 16823

**Restek France** Phone: (33) 01 60 78 32 10 • Fax: (33) 01 60 78 70 90 • restekfr@club-internet.fr

**Restek GmbH** Phone: (49) 06172 2797 0 • Fax: (49) 06172 2797 77 • RESTEK-GMBH@t-online.de

**Restek Ireland** Phone: (44) 28 9081 4576 • Fax: (44) 28 9081 4576 • restekeurope@aol.com

**Thames Restek UK Ltd** Phone: (44) 01494 563377 • Fax: (44) 01494 564990 • Sales@Thamesrestek.co.uk

Please direct your comments on this publication to Carrie Sprout, Graphic Designer, at carrie.sprout@restekcorp.com or call 814-353-1300, ext. 2151.



**Restek Corporation**  
110 Benner Circle  
Bellefonte, PA 16823-8812

Return Service Requested